

Certificate of Conformance

Page 1 of 1

Instrument: OL 730CV Programmable Radiometer/Photometer

Serial Number: 11203051

Instrument received within tolerance: No Inoperable Yes New
(If No or Inoperable, refer to repair report.)

Project No: 914-396

Calibration Date: 24-Oct-2011

Certification Date: 28-Oct-2011

Calibration Due: 28-Oct-2012

Calibration Equipment Used:

C/N: 000552 Keithley 263 Calibrator/Source SN: 0682726 Calibration Date: 09-May-2011 Calibration Due: 09-May-2012

C/N: 000265 Keithley 197 DMM SN: 431159 Calibration Date: 30-Sep-2011 Calibration Due: 30-Mar-2012

Environment Conditions:

Ambient Temperature: 23 °C ± 1°C

Relative Humidity: < 60 %

This is to certify the OL 730CV Programmable Radiometer/Photometer listed above, after inspection, testing, and/or service, has been found to be operating within Manufacturer's stated specifications.

Calibration was performed using standards and instrumentation traceable to the National Institute of Standards and Technology.

This certificate shall not be reproduced, except in full, without written approval of Gooch & Housego.

Certified by Gooch & Housego



Alex Mena / Jc. Electronic Technician

Original: \\HOST\REPORTS\CERTIFICATE\ECEQUIP\E000007.730C.730CV.730CMUX.XLT REV F 100203

DECLARATION OF CONFORMITY

(In accordance with EN 45014)

We,

***Optronic Laboratories, LLC
DBA Gooch & Housego
4632 36th Street
Orlando, Florida 32811 USA***

declare under our sole responsibility that the product,

Type of Equipment: Programmable Radiometer/Photometer
Model Number: OL 730C/ OL 730CV

to which this declaration relates is in conformity with the following standards or other normative documents:

EMC: EN 55011: 1991 Class A
EN 50082-1 1992
IEC 801-2 1991, 4 kV CD; 8 kV AD
IEC 801-3 1984, 3 V/m, CW
IEC 801-4 1988, 500 V Signal Lines
1 kV AC Power Lines
LVD: EN 61010-1: 1995

following the provision of EMC Directive 89/336/EEC and Low Voltage Directive 73/23/EEC.

Orlando, Florida 32811 U.S.A.

 9/23/97
Signature and Date

Bob Bronson, VP of Engineering
Typed Name and Title